

# TYPE APPROVAL CERTIFICATE

**This is to certify:****That the Misc. detector**with type designation(s)  
**MMT330, HMT330, DMT340, PTU300**

Issued to

**VAISALA OYJ**  
**Vantaa, Finland**is found to comply with  
**DNV GL rules for classification – Ships, offshore units, and high speed and light craft****Application :****Product(s) approved by this certificate is/are accepted for installation on all vessels classed by DNV GL.****Location classes:**

Type	Temperature	Humidity	Vibration	EMC	Enclosure
<b>MMT330</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B / IP65</b>
<b>HMT330</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B / IP65</b>
<b>DMT340</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B / IP65</b>
<b>PTU300</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B</b>	<b>B / IP65</b>

Issued at **Høvik** on **2017-02-08**This Certificate is valid until **2017-12-31**.DNV GL local station: **Helsinki**for **DNV GL**Approval Engineer: **Ståle Sneen**

---

**Odd Magne Nesvåg**  
**Head of Section**

This Certificate is subject to terms and conditions overleaf. Any significant change in design or construction may render this Certificate invalid. The validity date relates to the Type Approval Certificate and not to the approval of equipment/systems installed.



## Product description

The following types are covered by this approval:

- MMT330 Moisture and temperature transmitter for oil
- HMT330 Humidity and temperature transmitter
- DMT340 Dewpoint transmitter
- PTU300 Combined pressure, humidity and temperature transmitter

## Application/Limitation

The Type Approval covers hardware listed under Product description. When the hardware is used in applications to be classed by DNV GL, documentation for the actual application is to be submitted for approval by the manufacturer of the application system in each case. Reference is made to DNV GL rules for classification of ships Pt.4 Ch.9 Control and monitoring systems.

Type Approval of sensor is only valid for units with a valid calibration certificate.

## Type Approval documentation

### Tests carried out

Applicable tests according to Standard for Certification No. 2.4, April 2006.

### Marking of product

The products to be marked with:

- manufacturer name
- model name
- serial number
- power supply ratings

### Periodical assessment

The scope of the periodical assessment is to verify that the conditions stipulated for the type are complied with, and that no alterations are made to the product design or choice of systems, software versions, components and/or materials.

The main elements of the assessment are:

- Ensure that type approved documentation is available
- Inspection of factory samples, selected at random from the production line (where practicable)
- Review of production and inspection routines, including test records from product sample tests and control routines
- Ensuring that systems, software versions, components and/or materials used comply with type approved documents and/or referenced system, software, component and material specifications
- Review of possible changes in design of systems, software versions, components, materials and/or performance, and make sure that such changes do not affect the type approval given
- Ensuring traceability between manufacturer's product type marking and the type approval certificate

Periodical assessment is to be performed after 2 years and after 3.5 years. A renewal assessment will be performed at renewal of the certificate.

END OF CERTIFICATE